





# Enteral Colonic Stent




- A.** SEMS placement for occlusion due to colorectal cancer of the descending colon
- B.** Improvement of the patient's symptoms after stent placement

by Roberto Di Mitri et al [Scientific World Journal. 2014 Jan 2;2014:651765]

# Niti-S™ Enteral Colonic Stent

for malignant colorectal obstruction



## Features

- Unfixed cell with weaving construction
  - Maximize conformability to prevent migration and the risk of perforation
  - Optimal combination of radial and axial force to maintain patency in tortuous anatomy
- Radiopaque marker: 4 (four) at both ends & 2 (two) in the middle

## Ordering Information

TTS (Through The Scope)					OTW (Over The Wire)				
Code	Stent		Delivery		Code	Stent		Delivery	
	Diameter (mm)	Length (cm)	Profile (Fr)	Usable Length (cm)		Diameter (mm)	Length (cm)	Profile (Fr)	Usable Length (cm)
CDT**__	18, 20, 22, 24	6, 8, 10, 12, 14, 15	10	220	CD**__	18, 20, 22, 24	6, 8, 10, 12, 14, 15	16	70
	26, 28		10.5			26, 28, 30		18	

\* Available 120cm delivery system [Coding: CDT\*\*\_\_-12]

## Released Article

- \* Endoscopic stenting as bridge-to-surgery(BTS) in left-sided obstructing colorectal cancer: Experience with conformable stents by Andrea Parodi et al [Clin Res Hepatol Gastroenterol. 2016 Nov;40(5):638-644]
- \* Endoscopic Double Metallic Stenting in the Afferent and Efferent Loops for Malignant Afferent Loop Obstruction with Billroth II Anatomy by Nakahara K et al [Clin Endosc. 2016 Jan;49(1):97-9]
- \* The New Nitinol Conformable Self-Expandable Metal Stents for Malignant Colonic Obstruction: A Pilot Experience as Bridge to Surgery Treatment by Roberto DiMitti and Filippo Mocchiari [Scientific World Journal. 2014 Jan 2;2014:651765]
- \* Self-Expandable Metal Stent for Malignant Colonic Obstruction: Outcome in Proximal vs. Left Sided Tumor Localization by A. O. Tal et al [Z Gastroenterol. 2013 Jun;51(6):551-7]
- \* Novel method of stent insertion for malignant lower rectal obstruction with proximal releasing delivery system (with video) by Kee Myung Lee, MD, PhD et al [Gastrointest Endosc. 2013 Dec;78(6):930-3]
- \* Outcome and safety of self-expandable metallic stents for malignant colon obstruction: a Korean multicenter randomized prospective study by Dae Young Cheung et al [Surg Endosc. 2012 Nov;26(11):3106-13]
- \* Comparison of a Newly Designed Double-Layered Combination Covered Stent and D-Weave Uncovered Stent for Decompression of Obstructive Colorectal Cancer: A Prospective Multicenter Study by Chang Mo Moon, M.D. et al [Dis Colon Rectum. 2010 Aug;53(8):1190-6]